ORDER INFORMATION

CODE: DL1501 - 1 x 15ML + 1 x 5ML $DL1502 - 2 \times 15ML + 2 \times 5ML$ DL1503 - 4 x 30ML + 4 x 10ML DL1505 - 2 x 30ML + 2 x 10ML

DELTA **HDL - CHOLESTEROL** DIRECT ENZYMATIC CHOLORIMETRIC

INTENDED USE:

This reagent kit is intended for "in vitro" quantitative determination of High - Density Lipoprotein CholesteroL (HDL-C) in serum &

CLINICAL SIGNIFICANCE:

HDL particles serve to transport in the blood-stream.

HDL is known as "good cholesterol" because high levels are thought to lower the risk of heart disease and coronary artery disease. Alow HDL cholesterol level, is considered a greater heart disease risk

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

PRINCIPLE:

Direct determination of serum HDL (high-density lipoprotein cholesterol) levels without the need for any pre-treatment or centrifugation of the sample.

The method depends on the properties of a detergent which solubillizes only the HDL so that HDL-c is released to react with the cholesterol esterase, cholesterol oxidase and chromogens to give colour. The non HDL lipoprotein LDL, VLDL and chylomicrons are inhibited from reacting with the enzymes due to absorption of the detergents on their surfaces. The intensity of the color formed is proportional to the HDL concentration in the sample.

REAGENT COMPOSITION:

Reagent 1: Buffer Reagent Reagent 2: Substrate Reagent

HDL-C Calibrator: Concentration Printed on the Vial.

SAMPLES:

Serum or heparinized plasma, free of hemolysis, removed from the blood clot as soon as possible.

Anticoagulants containing citrate should not be used.

WORKING REAGENT PREPARATION & STABILITY:

Reagent 1 & Reagent 2 are Ready to use and are stable up to the expiry date stated on the label.

HDL-C Calibrator: Dissolve with distilled water (Qty Printed on the vial). Cap and mix gently to dissolve contents. Reconstituted calibrator is stable for 7 days at 2° to 8°C or 21 days at -20°.

GENERAL SYSTEM PARAMETERS:

Reaction type **Endpoint** Wave length 546 nm Light Path 1Cm Reaction Temperature 37°C Blank / Zero Setting Reagent Reagent Volume 1 ml Sample Volume $10 \mu l$ Incubation Time 10 Min. Calibrator Concentration Printed on vial Low Normal at 37°C 35 mg/dl High Normal at 37°C 88 mg/dl Linearity 200 mg/dl

ASSAY PROCEDURE:

	Blank	Calibrator	Sample
Reagent 1	750 μ l	750 μl	750 μ l
Calibrator		10 μl	
Sample			10 μΙ

Mix and incubate for 5 Minutes at 37°C and then add

Reagent 2	250 μl	250 μl	250 μl
-----------	---------------	---------------	---------------

Mix and read the optical density (A) after 10 minutes incubation at 37°C.

CALCULATION:

$$\mathsf{HDL\text{-}C} \;\; (\mathsf{mg}/\mathsf{dI}) = \frac{\mathsf{OD} \; \mathsf{of} \; \mathsf{Sample}}{\mathsf{OD} \; \mathsf{of} \; \mathsf{Calibrator}} \mathsf{X} \; \mathsf{Conc.} \; \mathsf{of} \; \mathsf{Calibrator}$$

LINEARITY:

Reagent is Linear up to 200 mg/dl.

Dilute the sample appropriately and re-assay if HDL-C Concentration exceeds 200 mg/l.

REFERENCE NORMAL VALUE:

Adult male: 35-79.5 mg/dl 42 - 88 mg/dl Adult female:

It is recommended that each laboratory should assign its own normal range.

QUALITY CONTROL:

For accuracy it is necessary to run known controls with every assay.

LIMITATION & PRECAUTIONS:

- 1. Storage conditions as mentioned on the kit to be adhered.
- Do not freeze or expose the reagents to higher temperature as it may affect the performance of the kit.
- Before the assay bring all the reagents to room temperature.
- 4. Avoid contamination of the reagent during assay process.
- Use clean glassware free from dust or debris.

BIBLIOGRAPHY:

- 1. Natio H KCholesterol Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeeton 1984; 1207-1213 and
- 2. US National Cholestrol Educatiopn Program of the National Institutes of Health.
- 3. Young DS. Effects of Drugs on Clinical Lab. Tests, 4th ad AACC Press, 1995.
- 4. Young DS. Effects of diseases on Clinical Lab. Tests 4th ad
- 5. Burlis A et al. Tietz Texbook of Clinical Chemistry, 3rd ed AACC 1999.
- 6. Tietz N W et al, Clinical to Laboratory Tests, 3rd ed AACC

Tel / Fax: 0 23 62 - 23 20 30, Email: delta@deltalab.in, Web: www.deltalab.in